




**Example 1.** PNL measurement on the CC must be within 1 cm of PNL measurement on the MLO.










**Example 2.** Measurement of the PNL on the CC


- Measurement is made from the base of the nipple to the edge of the image (if pec muscle is not visualized)
- Measurement is made from the base of the nipple to the anterior margin of the pectoralis muscle (if visualized)

B)













### **Example 3.** Measurement of the PNL on the MLO

• Measure from the base of the nipple to the anterior margin of the pectoralis muscle where the intersection can be measured as a 90° angle.





#### **CORRECT METHOD FOR DETERMINING PNL ANGLE - 90°**





#### **INCORRECT METHOD FOR DETERMINING PNL ANGLE - TOO STEEP: 110°**





**INCORRECT METHOD FOR DETERMINING PNL ANGLE - NOT STEEP ENOUGH: 80°** 

